

Operating Instructions

according to GUV 0.1
Area and Activities

AREA: **MFZ 3.034**

RESPONSIBLE: **IMCES**

INSTRUMENT

Zeiss AxioObserver 7 and Apotome.2

HARARDS FOR PEOPLE AND ENVIRONMENT



- Pinching hazard at mechanically controlled parts
- Hazard from genetically engineered and/or infectious organisms
- Possibly hazard from heat and CO₂ in the incubation chamber

PROTECTIVE MEASURES AND RULES



- Follow the operating instructions!
- Only use the device after thorough training
- Do not use the device in case of obvious damage
- Use personal protective equipment

IN CASE OF EMERGENCY

- Switch off the device immediately
- In case of injuries: Perform first aid as required

FIRST AID IN CASE OF ACCIDENTS

EMERGENCY TEL: 112

- Use the first aid kit in room 3.030 to treat wounds.
- Contact first aider for help: Prof. Winfried Siffert Tel: 0 201 723 3470
- Inform project leader about accident: Dr. Mike Hasenberg 0 201 723 82365



In emergencies:

- **Company doctor: Tel: 3480/3481**
Dr. med Beate Anna Streich
- **Emergency team at children's hospital and reanimation team: Tel: 1001**

In case of fire:

- **Call 112**
- **Use emergency showers if needed**
Leave building or use fire extinguisher (next to door)

INSPECTION - MAINTENANCE

- The device is under regular maintenance by the manufacturer company

Translation provided for your convenience. In case of contradictions,
the German version is to be considered authoritative.

Datum

31.08.2026